

Swiss Confederation

SCS Directory

Accreditation number: SCS 0014

International standard: ISO/IEC 17025:2017
 Swiss standard: SN EN ISO/IEC 17025:2018

MOSER TMT AG
 Temperatur-Messtechnik
 Sommeraustasse 2
 8634 Hombrechtikon

Head: Michael Moser
 Responsible for MS: Michael Moser
 Telephone: +41 55 244 23 46
 E-Mail: info@moserag.ch
 Internet: <http://www.moserag.ch>
 Initial accreditation: 01.03.1988
 Current accreditation: 18.03.2024 to 17.03.2029
 Scope of accreditation see: www.sas.admin.ch
 (Accredited bodies)

Scope of accreditation as of 18.03.2024

Calibration laboratory for the quantity temperature

Calibration and Measurement Capability (CMC)

Measured Quantity / Instrument or Gauge	Measurement Range	Measurement Conditions	Best Measurement Uncertainty \pm ¹⁾	Remarks
Temperature		Calibration bath		Comparison with standard platinum-resistance thermometers
Direct contact thermometer with display	-80 °C ... 40 °C > 40 °C ... 150 °C > 150 °C ... 420 °C		0,020 K 0,025 K 0,030 K	
Contact thermometer Resistance thermometer				
Fixed-point for ITS-90	0.01 °C	Tripel point of water	0,010 K	

In case of contradictions in the language versions of the directories, the German version shall apply.

* / * / * / * / *