



## SCS Directory

**Accreditation number: SCS 0114**

International standard: ISO/IEC 17025:2017  
Swiss standard: SN EN ISO/IEC 17025:2018

AMETEK CTS GmbH  
Sternenhofstrasse 15  
4153 Reinach  
Switzerland

Head: Erich Brogli  
Responsible for MS: Sven Schäuble  
Telephone: +41 61 204 41 11  
E-Mail: [erich.brogli@ametek.com](mailto:erich.brogli@ametek.com)  
Internet: <http://www.ametek-cts.com>  
Initial accreditation: 22.06.2009  
Current accreditation: 22.06.2019 to 21.06.2024  
Scope of accreditation see: [www.sas.admin.ch](http://www.sas.admin.ch)  
(Accredited bodies)

Site under the accreditation:

Lünener Strasse 211  
59174 Kamen  
(Germany)

Head: Michael Kreisel  
Responsible for MS: Manuel Mihlan  
Telephone: +49 2307 26070 0  
E-Mail: [michael.kreisel@ametek.com](mailto:michael.kreisel@ametek.com)  
Internet: <http://www.ametek-cts.com>

## Scope of accreditation as of 06.04.2023

### Calibration laboratory for electrical quantities

#### Calibration and Measurement Capability (CMC)

Measured Quantity / Instrument or Gauge	Measurement Range	Measurement Conditions	Best Measurement uncertainty $\pm$ <sup>1)</sup>	Remarks
<b>DC Voltage</b> <sup>2)</sup>	1 mV ... 1000 V		0,20 %	Output Voltage of Arbitrary Generators and Voltage Sources ISO 7637-2
<b>DC High Voltage</b> <sup>2)</sup>	100 V ... 30 kV		3,18 %	Output Voltage of Test Generators and Charging Voltages of Pulse Circuits IEC 61000-4-x ISO 7637-2

<sup>1)</sup>The given extended measurement uncertainty is the standard uncertainty of the measurement multiplied by an extension factor  $k = 2$ , which corresponds to a confidence level of about 95% for a normal distribution.

<sup>2)</sup> On-site calibrations with increased measurement uncertainty possible.



## SCS Directory

## Accreditation number: SCS 0114

Measured Quantity / Instrument or Gauge	Measurement Range	Measurement Conditions	Best Measurement uncertainty $\pm$ <sup>1)</sup>	Remarks
<b>AC Voltage</b> <sup>2)</sup>	1 mV ... 1000 V		0,23 %	Arbitrary Generators, Power Fail Generators, Variacs IEC 61000-4-x ISO 7637-2
<b>DC Current</b> <sup>2)</sup>	1 mA ... 100 A		0,20 %	
<b>AC Current</b> <sup>2)</sup>	1 mA ... 100 A		0,20 %	
<b>Frequency, Sine</b> <sup>2)</sup>	DC ... 1 MHz 10 kHz ... 4 GHz		4·10 <sup>-3</sup> 17,9·10 <sup>-6</sup>	Only DC ... 1 MHz on-site possible
<b>DC Resistance</b> <sup>2)</sup>	1 m $\Omega$ ... 10 M $\Omega$		0,24 %	
<b>Inductance</b> <sup>2)</sup>	1 $\mu$ H ... 1000 H	f=1 kHz	0,21 %	
<b>Capacitance</b> <sup>2)</sup>	100 pF ... 100 $\mu$ F	f=1 kHz	0,44 %	
<b>ESD</b>				IEC 61000-4-2 ISO 10605
Voltage	100 V ... 30 kV	DC	0,17 %	
Current Peak	0 A ... 120 A		4,70 %	
Current	Current	@30 ns @60 ns @60-800 ns	4,70 % + (12 %)* 4,70 % + (8 %)* 4,70 % + (8 %)* (%)* Reproducibility device setup	
Rise Time	500 ps ... 1 $\mu$ s		30,47 ps	
<b>Burst into 50 <math>\Omega</math></b> <sup>2)</sup>				IEC 61000-4-4 ISO 7637-2
Voltage	20 V ... 6000 V 200 V ... 8000 V 20 V ... 8000 V	Common Mode	4,95 % 5,01 % 7,63 %	
Rise Time	1 ns ... 1 $\mu$ s		184 ps	
Pulse Duration	10 ns ... 10 $\mu$ s		1602 ps	
Repetition Frequency	1 Hz ... 1 MHz		4·10 <sup>-3</sup>	
<b>Burst into 1000 <math>\Omega</math></b> <sup>2)</sup>				IEC 61000-4-4 ISO 7637-2
Voltage	20 V ... 6000 V 200 V ... 8000 V 25 V ... 1000 V	Common Mode	5,04 % 5,35 % 7,70 %	
Rise Time	1 ns ... 1 $\mu$ s		258 ps	

<sup>1)</sup>The given extended measurement uncertainty is the standard uncertainty of the measurement multiplied by an extension factor k = 2, which corresponds to a confidence level of about 95% for a normal distribution.

<sup>2)</sup> On-site calibrations with increased measurement uncertainty possible.



## SCS Directory

## Accreditation number: SCS 0114

Measured Quantity / Instrument or Gauge	Measurement Range	Measurement Conditions	Best Measurement uncertainty $\pm$ <sup>1)</sup>	Remarks
Pulse Duration	10 ns ... 10 $\mu$ s		2,4 ns	ISO 7637-2 IEC 61000-4-5
Repetition Frequency	1 Hz ... 1 MHz		$4 \cdot 10^{-3}$	
<b>Pulses <math>\mu</math>s range <sup>2)</sup></b>				
Voltage	100 V ... 12000 V 1000 V ... 20000V		4,87 % 4,96 %	
Rise Time	100 ns ... 100 $\mu$ s	Voltage	2,1 ns	ISO 7637-2
		Current	51 ns	
Pulse Duration	1 $\mu$ s ... 1000 $\mu$ s	Voltage	540 ns	
		Current	128 ns	
Current	1 A ... 1000 A 100 A ...20'000 A		3,67 % 3,88 %	ANSI C62.41 IEC 61000-4-12
<b>Pulses ms range <sup>2)</sup></b>				
Voltage	10 V ... 1000 V		4,65 %	
Rise Time	50 $\mu$ s ... 50 ms		3,1 ns	
Pulse Duration	1000 $\mu$ s ... 1500 ms		2,6 $\mu$ s	IEC 61000-4-12
Current	1 A ... 5000 A		6,46 %	
<b>Ringwave <sup>2)</sup></b>				
Voltage	200 V ... 6000 V		3,94 %	
Current	5,3 A ... 500 A		1,81 %	Not requested by -12
Rise Time	0,1 $\mu$ s ... 5 $\mu$ s	Voltage	2,26 ns	
		Current	2.01 ns	
Frequency	DC ... 250 kHz		$4 \cdot 10^{-3}$	
<b>Damped Oscillatory <sup>2)</sup></b>				Not requested by -12
Voltage	200 V ... 4000V		1.13 %	
Current	1 A ... 120A		1,81 %	
Rise Time	0,05 $\mu$ s ... 5 $\mu$ s	Voltage	10.22 ns	
		Current	51 ns	
Frequency	DC ... 1 MHz		$4 \cdot 10^{-3}$	

<sup>1)</sup>The given extended measurement uncertainty is the standard uncertainty of the measurement multiplied by an extension factor  $k = 2$ , which corresponds to a confidence level of about 95% for a normal distribution.

<sup>2)</sup> On-site calibrations with increased measurement uncertainty possible.



## SCS Directory

## Accreditation number: SCS 0114

Measured Quantity / Instrument or Gauge	Measurement Range	Measurement Conditions	Best Measurement uncertainty $\pm$ <sup>1)</sup>	Remarks
<b>Fast Damped Oscillator</b> <sup>2)</sup>				IEC 61000-4-18
Voltage	200 V ... 6000V		3,57 %	
Current	1 A ... 120A		3,79 %	
Rise Time	0,5 ns ... 1 $\mu$ s	Voltage	242,77 ps	
	0,5 ns ... 10 $\mu$ s	Current	231,17 ps	
Frequency	DC ... 100 MHz		4 · 10 <sup>-3</sup> %	
<b>Power Fail</b> <sup>2)</sup>				IEC 61000-4-11
Voltage (cont.)	-400 V ... 400 V	With or without load	0,05 %	
Peak Voltage (Over- and Undershoot)	-20 V ... 270 V	100 $\Omega$ load	3,35 %	
Rise Time/Fall Time	0,1 $\mu$ s ... 10 $\mu$ s	100 $\Omega$ load	1,78 ns	
Current	10 A ... 50 A	100 $\Omega$ load	2,27 %	
Peak Current (Inrush)	200 A ... 1000 A		1,94 %	
Phase	0 ° ... 360 °		1,27 °	
<b>ESD target</b>				IEC 61000-4-2 ISO 10605
Input Impedance	0,1 $\Omega$ ... 50 $\Omega$	DC	1,36 %	
Transfer Impedance	0,1 $\Omega$ ... 50 $\Omega$	DC	1,38 %	
Insertion Loss	20 dB ... 60 dB	20 kHz ... 4 GHz	0,66 dB	
<b>ESD target adapter</b>				IEC 61000-4-2 ISO 10605
Return Loss (low reflect)	-60dB ... -20dB	20 kHz ... 1 GHz	0,03	
	-40dB ... -5dB	>1 GHz ... 4 GHz	0,04	
Insertion Loss	0 dB ... 10dB	20 kHz ... 4 GHz	0,07 dB	
<b>Burst adapter (High Imp)</b>				IEC 61000-4-4
Input Impedance	100 $\Omega$ ... 100 k $\Omega$	DC	0,16 %	
Insertion loss	45 dB ... 65dB	20 kHz ... 400 MHz	0,61 dB	
<b>Burst adapter (Match)</b>				IEC 61000-4-4
Input Impedance	40 $\Omega$ ... 60 $\Omega$	DC	0,16 %	
Insertion loss	35 dB ... 55 dB	20 kHz ... 400 MHz	0,39 dB	

<sup>1)</sup>The given extended measurement uncertainty is the standard uncertainty of the measurement multiplied by an extension factor k = 2, which corresponds to a confidence level of about 95% for a normal distribution.

<sup>2)</sup> On-site calibrations with increased measurement uncertainty possible.



## SCS Directory

## Accreditation number: SCS 0114

Measured Quantity / Instrument or Gauge	Measurement Range	Measurement Conditions	Best Measurement uncertainty $\pm$ <sup>1)</sup>	Remarks
<b>RF-voltage (50 <math>\Omega</math>)</b>	200 $\mu$ V ... 10 V	9 kHz ... 3 GHz	1,4 %	IEC 61000-4-6
	0,01 V ... 100 V	100 kHz ... 2 GHz	1,5 %	
<b>Amplitude Modulation</b>				IEC 61000-4-6
Frequency	100 Hz ... 10 MHz	Carrier f = 100 kHz ... 300 MHz	18·10 <sup>-6</sup>	
Modulation Index	10 % ... 95 %	Carrier f = 100 kHz ... 300 MHz	1,1 %	
<b>Spurious emissions</b>	0 dBc ... 50 dBc	Carrier f = 100 kHz ... 300 MHz	0,69 dB	IEC 61000-4-6
<b>150-50<math>\Omega</math>-Adapter</b>				IEC 61000-4-6
Insertion Loss	5 dB ... 12 dB	20 kHz ... 300 MHz	0,29 dB	
	5 dB ... 12 dB	>300 MHz...1 GHz	0,47 dB	
<b>Matching attenuator</b>				Coaxial Connector
Attenuation	0 dB ... 40 dB	20 kHz ... 300 MHz	0,12 dB	
	0 dB ... 40 dB	>300 MHz...1 GHz	0,21 dB	
<b>CDN</b>				IEC 61000-4-6
Output Impedance	90 $\Omega$ ... 210 $\Omega$	20 kHz ... 300 MHz	8,4 $\Omega$	
<b>50 Ohm load</b>				IEC 61000-4-6
Impedance	40 $\Omega$ ... 60 $\Omega$	20 kHz ... 300 MHz	1,6 $\Omega$	
		>300 MHz ... 1 GHz	2,9 $\Omega$	
S11	$\leq$ 0.2	20 kHz ... 1 GHz	0,02	
<b>AC – DC – Harmonics Source</b> <sup>2)</sup>				EN/IEC 61000-4-11, EN/IEC 61000-4-13, EN/IEC 61000-4-14, EN/IEC 61000-4-17, EN/IEC 61000-4-28, EN/IEC 61000-4-29, EN/IEC 61000-3-2, EN/IEC 61000-3-3, EN/IEC 61000-3-11, EN/IEC 61000-3-12
AC Voltage RMS	1 mV ... 750 V	15 Hz ... 850 Hz	0,2 %	
AC Voltage RMS	1 mV ... 100 V	1 Hz ... 10 kHz	0,2 %	
AC Voltage Peak	1 mV ... 1000 V	15 Hz ... 850 Hz	0,2 %	
Noise on DC	1 mV ... 10 V	1 Hz ... 20 MHz	1,08 %	
DC Voltage	1 mV ... 1000 V		0,1 %	

<sup>1)</sup>The given extended measurement uncertainty is the standard uncertainty of the measurement multiplied by an extension factor k = 2, which corresponds to a confidence level of about 95% for a normal distribution.

<sup>2)</sup> On-site calibrations with increased measurement uncertainty possible.



## SCS Directory

## Accreditation number: SCS 0114

Measured Quantity / Instrument or Gauge	Measurement Range	Measurement Conditions	Best Measurement uncertainty $\pm$ <sup>1)</sup>	Remarks
AC Current RMS	0,1 A ... 200 A	15 Hz ... 500 Hz	0,3 %	EN/IEC 61000-4-7, EN/IEC 61000-3-2, JIS C 61000-3-2, EN/IEC 61000-3-12
Rise- / Fall-Time DC, AC	100 ns ... 500 us	0 V ... 1000 V	< 2 ns	
Frequency	1 Hz ... 10 kHz		0,012 %	
THD Voltage	10 Hz ... 10 kHz		0,253 %	
Phase	0 ° ... 360 °	1 Hz ... 10 kHz	2 °	
<b>Harmonics &amp; Flicker Power Analyzer</b> <sup>2)</sup>				
AC Voltage RMS	1 V ... 1000 V	16 Hz ... 10 kHz	0,32 %	
AC Current RMS	2.2 A ... 80 A	10 Hz ... 1 kHz	0,18 %	
AC Current RMS	1 mA ... 2,2 A	10 Hz ... 5 kHz	< 0,35 %	
THD Voltage	0 % ... 10 %	16 Hz ... 6kHz	0,32 %	
THD Current	0 % ... 10 %	16 Hz ... 6kHz	0,34 %	
<b>Flicker</b> <sup>2)</sup>				IEC 61000-4-15, IEC 61000-3-3, IEC 61000-3-11
PST	0 ... 5	Set values of CPM, dV/V as per standard	0,124 %	
dmax	0 % ... 10 %	Use of procedure as per standard	1,0 %	
dc	0 % ... 10 %	Use of procedure as per standard	1,0 %	
Tmax	0 s ... 10s	Use of procedure as per standard	1,0 %	
<b>Flicker Impedance</b> <sup>2)</sup>				IEC 61000-3-3, -11, IEC TR 60725
Flicker Impedance R	0,1 R ... 0,5 R		< 0,118 %	
Flicker Impedance L	100 uH ... 1 mH		< 0,118 %	
Flicker Impedance Z	0,1 R ... 0,5 R		0,118 %	ISO 7637-2, ISO 11452-4, CISPR 16-1-2, CISPR 25, GS 95002
<b>Artificial Network</b> <sup>2)</sup>				

<sup>1)</sup>The given extended measurement uncertainty is the standard uncertainty of the measurement multiplied by an extension factor  $k = 2$ , which corresponds to a confidence level of about 95% for a normal distribution.

<sup>2)</sup> On-site calibrations with increased measurement uncertainty possible.



## SCS Directory

## Accreditation number: SCS 0114

Measured Quantity / Instrument or Gauge	Measurement Range	Measurement Conditions	Best Measurement uncertainty $\pm$ <sup>1)</sup>	Remarks
Impedance Absolute	0,5 $\Omega$ ... 110 $\Omega$	0,1 ... 0,5 MHz	0,3 $\Omega$	
	0,5 $\Omega$ ... 110 $\Omega$	0,5 ... 1.0 MHz	0,3 $\Omega$	
	0,5 $\Omega$ ... 110 $\Omega$	1,0 ... 15 MHz	0.7 $\Omega$	
	0,5 $\Omega$ ... 110 $\Omega$	15 ... 100 MHz	0.7 $\Omega$	
	0,5 $\Omega$ ... 110 $\Omega$	100 ... 200 MHz	0.7 $\Omega$	
Impedance Phase	0,1° ... 120°	0,1 ... 0,3 MHz	3.5 °	
	0,1° ... 120°	0,3 ... 10 MHz	1.2 °	
	0,1° ... 120°	10 ... 120 MHz	0.8 °	
	0,1° ... 120°	120 ... 200 MHz	0.9 °	
Isolation	0.0 ... 70 dB	0,1 ... 1.0 MHz	0,5 dB	
	0.0 ... 70 dB	1.0 ... 100 MHz	1.0 dB	
	0.0 ... 70 dB	100 ... 200 MHz	0.6 dB	
VDF	0.0 ... 30 dB	0,1 ... 200 MHz	0,5 dB	

All quantities: in/from<sup>2)</sup> both sites.

\* / \* / \* / \* / \*

<sup>1)</sup>The given extended measurement uncertainty is the standard uncertainty of the measurement multiplied by an extension factor  $k = 2$ , which corresponds to a confidence level of about 95% for a normal distribution.

<sup>2)</sup> On-site calibrations with increased measurement uncertainty possible.