



## SCS Directory

## Accreditation number: SCS 0149

International standard: ISO/IEC 17025:2017  
Swiss standard: SN EN ISO/IEC 17025:2018

Martin & Tschopp GmbH  
Calibration Laboratory SCS  
Hauptstrasse 91  
4434 Hölstein

Head: Peter Grahle  
Responsible for MS: Frank Möller  
Telephone: +41 61 956 17 00  
E-Mail: [info@mt-swiss.com](mailto:info@mt-swiss.com)  
Internet: <http://www.mt-swiss.com>  
Initial accreditation: 05.10.1994  
Current accreditation: 01.01.2022 to 31.12.2026  
Scope of accreditation see: [www.sas.admin.ch](http://www.sas.admin.ch)  
(Accredited bodies)

### Scope of accreditation as of 01.01.2022

#### Calibration laboratory for length and roundness

##### Calibration and Measurement Capability (CMC)

Measured Quantity / Instrument or Gauge	Measurement Range	Measurement Conditions	Best Measurement Uncertainty $\pm$ <sup>1)</sup>	Remarks
<b>LENGTH</b>  measurements of one-dimensional gauges	external dimensions  to 20 mm		0,30 $\mu$ m	on one Coordinate Measuring Machine
<b>ROUNDNESS</b> Cylindrical gauges	external $\varnothing$ 1 mm ... 20 mm	filter 15 W/U	0,16 $\mu$ m	with Roundness measuring instrument
Roundness	$\varnothing$ 0.05 mm ... 20 mm	measurements with three- points method of ISO 4292:1985	0.15 $\mu$ m + 0.14•ROM ROM : Roundness	