



## SCS Directory

Accreditation number: SCS 0157

International standard: ISO/IEC 17025:2017  
Swiss standard: SN EN ISO/IEC 17025:2018

Berlinger & Co. AG  
Calibration Laboratory  
Toggenburgerstrasse 20  
9608 Ganterschwil

Head: Janik Baumgartner  
Responsible for MS: Daniel Steiner  
Telephone: +41 71 982 88 11  
E-Mail: [info@berlinger.com](mailto:info@berlinger.com)  
Internet: [www.berlinger.com](http://www.berlinger.com)  
Initial accreditation: 13.05.2020  
Current accreditation: 13.05.2025 to 12.05.2030  
Scope of accreditation see: [www.sas.admin.ch](http://www.sas.admin.ch)  
(Accredited bodies)

### Scope of accreditation as of 13.05.2025

#### Calibration laboratory for temperature

##### Calibration and Measurement Capability (CMC)

Measured Quantity / Instrument or Gauge	Measurement Range	Measurement Conditions	Best Measurement Uncertainty $\pm$ <sup>1)</sup>	Remarks
<b>Temperature</b> Calibration of thermometers with direct display and temperature data loggers with resistance sensors	-40 °C ... 85 °C	Calibration bath	0,080 K	Comparison with standard platinum resistance thermometers
	0 °C	Ice point (H <sub>2</sub> O)	0,024 K	

In case of contradictions in the language versions of the directories, the German version shall apply.

\* / \* / \* / \* / \*