



## SCS Directory

Accreditation number: SCS 0068

International standard: ISO/IEC 17025:2017  
Swiss standard: SN EN ISO/IEC 17025:2018

walter + bai ag  
Calibration laboratory  
Industriestrasse 4  
8224 Löhningen

Head: R. Hell  
Responsible for MS: M. Rieser  
Telephone: +41 52 687 25 26  
E-Mail: [roger.hell@walterbai.com](mailto:roger.hell@walterbai.com)  
Internet: <http://www.walterbai.com>  
Initial accreditation: 19.10.1995  
Current accreditation: 20.12.2020 to 19.12.2025  
Scope of accreditation see: [www.sas.admin.ch](http://www.sas.admin.ch)  
(Accredited bodies)

### Scope of accreditation as of 09.08.2022

#### Calibration laboratory for material testing machines (Measurand Force, Torque, Length, Angle, Impact, Pressure, Roughness and Hardness)

##### Calibration and Measurement Capability (CMC)

Measured Quantity / Instrument or Gauge	Measurement Range	Measurement Conditions	Best Measurement Uncertainty $\pm$ <sup>1)</sup>	Remarks
<b>LOAD</b> Tension and Compression	0.01 N ... < 2.00 N 2 N ... < 1000 N	With loading pieces	0.0002 N 0.0002 N	Calibration/Verification of Material testing machines according ISO 7500-1 / ASTM E4
	200 N ... 240 kN	With loadcell class 0.5 according EN ISO 376 and ASTM E74-00	0.06 %	
	20 N ... 200 N	With loadcell class 0.5 according EN ISO 376 and ASTM E74-00	0.12 %	



## SCS Directory

Accreditation number: SCS 0068

Measured Quantity / Instrument or Gauge	Measurement Range	Measurement Conditions	Best Measurement Uncertainty $\pm$ <sup>1)</sup>	Remarks		
Tension	20 kN ... 1500 kN >1500 kN ... 3 MN	With borrow equipment from SCS 0069	0.12 % 0.02 %	Calibration/Verification of Rotary bending testing machines according ISO 1143 / DIN 50113		
Compression	400 kN ... 5 MN >5 MN ... 10 MN		0.12 % 0.02%			
Bendingload	0.001 Nm...<0.8 Nm 0.25 Nm...<40 Nm	With loading pieces	0.00002 Nm 0.00002 Nm			
<b>PRESSURE</b>						
Static	0 bar ... 20 bar	DKD-R6-1 / ISO 7500-1 ASTM E4 / EN 837-1	0.3 %, but not smaller than 20 mbar		Pressure calibration of testing machines. Pressure sensors and Manometers	
	20 bar ... < 500 bar		0.2 %			DKD-R6-1 / EN 837-1 or in dependence ISO 7500-1
	500 bar ... 5000 bar		0.3 %			DKD-R6-1 / EN 837-1 or in dependence ISO 7500-1
<b>CALIBRATION OF IMPACT-TESTING-MACHINES</b>						
	15 J ... 950 J	ISO 148-2 /ASTM E23-96	Limits of deviation according the standard of the procedure		Only calibration / Verification according ISO 148-2	
<b>LENGTH</b>						
Extensometer	Till 50 mm	EN ISO 9513 and ASTM E83 18 °C < T < 28 °C		Clip-On or permanent installation		
	Resolution 0.1 $\mu$ m		0.6 $\mu$ m + 1•10 <sup>-4</sup> •L			
	Resolution 0.2 $\mu$ m		0.6 $\mu$ m + 1•10 <sup>-4</sup> •L			
	Resolution 0.5 $\mu$ m		0.8 $\mu$ m + 1•10 <sup>-4</sup> •L			
	Resolution 1.0 $\mu$ m		1.3 $\mu$ m + 1•10 <sup>-4</sup> •L			



## SCS Directory

## Accreditation number: SCS 0068

Measured Quantity / Instrument or Gauge	Measurement Range	Measurement Conditions	Best Measurement Uncertainty $\pm$ <sup>1)</sup>	Remarks
Displacement transducer	Till 60 mm	Reference KMF1		Plug free or spring loaded ISO-9513, ASTM E83
	Resolution 1 $\mu\text{m}$		$1.3 \mu\text{m} + 1 \cdot 10^{-4} \cdot L$	
	Resolution 2 $\mu\text{m}$		$2.4 \mu\text{m} + 1 \cdot 10^{-4} \cdot L$	
Displacement transducer or Extensometer	Resolution 10 $\mu\text{m}$	18 °C < T < 28 °C	$11.6 \mu\text{m} + 1 \cdot 10^{-4} \cdot L$	Plug free or spring loaded
	Till 400 mm			
	Resolution 0.1 $\mu\text{m}$		$0.7 \mu\text{m} + 7.1 \cdot 10^{-4} \cdot L$	
	Resolution 0.2 $\mu\text{m}$		$0.7 \mu\text{m} + 7.1 \cdot 10^{-4} \cdot L$	
Dial Gauge and Caliper	Resolution 0.5 $\mu\text{m}$	Reference KMF1	$0.9 \mu\text{m} + 7.1 \cdot 10^{-4} \cdot L$	Plug free or spring loaded
	Resolution 1 $\mu\text{m}$		$1.3 \mu\text{m} + 7.1 \cdot 10^{-4} \cdot L$	
	Till 50 mm			
	Resolution 1 $\mu\text{m}$		$2.2 \mu\text{m} + 1 \cdot 10^{-4} \cdot L$	
Piston- and Cross-Head-Measuring System	Resolution 2 $\mu\text{m}$	Newall, incremental	$3.0 \mu\text{m} + 1 \cdot 10^{-4} \cdot L$	Testing-Machines with installed displacement transducers
	Resolution 10 $\mu\text{m}$		$11.7 \mu\text{m} + 1 \cdot 10^{-4} \cdot L$	
	Resolution 20 $\mu\text{m}$		$23.2 \mu\text{m} + 1 \cdot 10^{-4} \cdot L$	
Piston- and Cross-Head-Measuring System	0 ... 500 mm	With steel-scale		Testing-Machines with installed displacement transducers
	Resolution 1 $\mu\text{m}$		$5 \mu\text{m} + 7 \cdot 10^{-4} \cdot L$	
	Resolution 2 $\mu\text{m}$		$7.5 \mu\text{m} + 7 \cdot 10^{-4} \cdot L$	
Piston- and Cross-Head-Measuring System	Resolution 10 $\mu\text{m}$	With steel-scale	$12.5 \mu\text{m} + 7 \cdot 10^{-4} \cdot L$	Testing-Machines with installed displacement transducers
	Till 1000 mm			
<b>TORSION</b>	Resolution 1 mm	With weights an lever loadingl	$250 \mu\text{m} + L$	Only for Torsion-Testing-Machines
	0.01 Nm ... 2.00 Nm		0.3 %	
	2 Nm ... 1000 Nm		0.15 %	
	5 Nm ... 240 Nm	Static procedure	0.15 %	
	5 Nm ... 240 Nm	Pseudo-static procedure	0.3 %	



**SCS Directory**

**Accreditation number: SCS 0068**

Measured Quantity / Instrument or Gauge	Measurement Range	Measurement Conditions	Best Measurement Uncertainty $\pm$ <sup>1)</sup>	Remarks
<b>ANGLE</b>	>240 Nm ... 6 kNm	With borrow equipment from SCS0069	0.16%	Only for Torsion-Testing-Machines
	0 Nm ... 100 kNm	With reference-load-cell and lever loading	0.3 %	
<b>ROUGHNESS RA</b>	0.2 $\mu$ m ... 12.5 $\mu$ m	18 °C <T< 28 °C	0.01 °	ISO-4287-T1 Only for testing of compression platen on testing machines
<b>CALIBRATION OF HARDNESS TESTING MACHINES</b>		Direct and Indirect testing procedure		
Hardness - Brinell	Procedure: HBW	ISO 6506-2 ASTM E10-10	Limits of deviation according the standard of the procedure	With reference piece according ISO 6506-3 ASTM E10-10
Hardness - Rockwell	Procedure: HRB, HRC	ISO 6508-2 ASTM E18-08	Limits of deviation according the standard of the procedure	With reference piece according ISO 6508-3 ASTM E18-08
Hardness - Vickers	Procedure: HV0.1, HV 0.2, HV0.3 HV1, HV5, HV10 HV20, HV 50, HV100	ISO 6507-2 ASTM E92-82/E384	Limits of deviation according the standard of the procedure	With reference piece according ISO 6507-3 ASTM E92-82/E384

All of the calibrations also are possible on customer side.

In case of contradictions in the language versions of the directories, the German version shall apply.

\* / \* / \* / \* / \*

1) The given extended measurement uncertainty is the standard uncertainty of the measurement multiplied by an extension factor k = 2, which corresponds to a confidence level of about 95% for a normal distribution.