



SCS Directory

Accreditation number: SCS 0171

International standard: ISO/IEC 17025:2017
Swiss standard: SN EN ISO/IEC 17025:2018

Proceq AG
Calibration Lab
Ringstrasse 2
8603 Schwerzenbach

Head: Bernardo Rieder
Responsible for MS: Joanna Lubowiecka
Telephone: +41 043 355 38 00
E-Mail: quality@screeningeagle.com
Internet: www.screeningeagle.com
Initial accreditation: 03.08.2022
Current accreditation: 03.08.2022 to 02.08.2027
Scope of accreditation see: www.sas.admin.ch
(Accredited bodies)

Scope of accreditation as of 28.11.2023

Calibration laboratory for hardness and optical measurements

Calibration and Measurement Capability (CMC)

Measured Quantity / Instrument or Gauge	Measurement Range	Measurement Conditions	Best Measurement Uncertainty \pm ¹⁾	Remarks
Hardness:				
Leeb	HLD 450 - 800 Leeb	Indirect	0,97 %	Acc. to ISO 16859-2
Hardness testing device	HLE 550 - 850 Leeb	Indirect	0,97 %	Acc. to ISO 16859-2
	HLG 400 - 600 Leeb	Indirect	0,97 %	Acc. to ISO 16859-2
	HLS 600 - 900 Leeb	Indirect	0,97 %	Acc. to ISO 16859-2
	HLC 500 - 850 Leeb	Indirect	0,97 %	Acc. to ISO 16859-2
	HLDL 650 - 900 Leeb	Indirect	0,97 %	Acc. to ISO 16859-2
Leeb hardness test blocks	HLD 450 - 800 Leeb	Indirect	0,95 %	Acc. to ISO 16895-3
	HLE 550 - 850 Leeb	Indirect	0,95 %	Acc. to ISO 16895-3
	HLG 400 - 600 Leeb	Indirect	0,95 %	Acc. to ISO 16895-3



SCS Directory

Accreditation number: SCS 0171

Measured Quantity / Instrument or Gauge	Measurement Range	Measurement Conditions	Best Measurement Uncertainty \pm ¹⁾	Remarks
Portable Vickers testing device	HLS 600 - 900 Leeb	Indirect	0,95 %	Acc. to ISO 16895-3
	HLC 500 - 850 Leeb	Indirect	0,95 %	Acc. to ISO 16895-3
	HLDL 650 - 900 Leeb	Indirect	0,95 %	Acc. to ISO 16895-3
	UCI HV1 / 200 - 900 HV	Indirect	1,5 * U _{crm}	Acc. to DIN 50159-2
	UCI HV5 / 200 - 900 HV	Indirect	1,5 * U _{crm}	Acc. to DIN 50159-2
	UCI HV10 / 200 - 900 HV	Indirect	1,5 * U _{crm}	Acc. to DIN 50159-2
Portable Rockwell testing device	HRC 50N 20-65 HRC	Indirect	0,5 HRC	Acc. to DIN 50157-2
Optical measures:				
Retroreflectometer Road marking				
Day visibility (Q _d)	5 mcd · m ⁻² · lx ⁻¹ ... 400 mcd · m ⁻² · lx ⁻¹	Indirect	10 %	Only Proceq / Zehntner device, Measuring geometry acc. to EN 1436:2018
Night visibility (R _L)	5 mcd · m ⁻² · lx ⁻¹ ... 2000 mcd · m ⁻² · lx ⁻¹	Indirect	10 %	Only Proceq / Zehntner device, Measuring geometry acc. to EN 1436:2018
Retroreflectometer traffic signs / foils				
Specific reflection value R _A (Night visibility)	1 cd · m ⁻² · lx ⁻¹ ... 2000 cd · m ⁻² · lx ⁻¹	Indirect	7 %	Only Proceq / Zehntner device, Measuring geometry acc. to ASTM E1709, ASTM E2540, CUAP 01.06/04, EN 67520, EN 12899-1, ISO 20471, ECE 104

In case of contradictions in the language versions of the directories, the German version shall apply.

* / * / * / * / *